imall

Chipsmall Limited consists of a professional team with an average of over 10 year of expertise in the distribution of electronic components. Based in Hongkong, we have already established firm and mutual-benefit business relationships with customers from, Europe, America and south Asia, supplying obsolete and hard-to-find components to meet their specific needs.

With the principle of "Quality Parts, Customers Priority, Honest Operation, and Considerate Service", our business mainly focus on the distribution of electronic components. Line cards we deal with include Microchip, ALPS, ROHM, Xilinx, Pulse, ON, Everlight and Freescale. Main products comprise IC, Modules, Potentiometer, IC Socket, Relay, Connector. Our parts cover such applications as commercial, industrial, and automotives areas.

We are looking forward to setting up business relationship with you and hope to provide you with the best service and solution. Let us make a better world for our industry!



Contact us

Tel: +86-755-8981 8866 Fax: +86-755-8427 6832 Email & Skype: info@chipsmall.com Web: www.chipsmall.com Address: A1208, Overseas Decoration Building, #122 Zhenhua RD., Futian, Shenzhen, China




Technical Data Sheet

Rosenberger

Calibration Adaptor RPC-2.92 Jack - SMP Jack

02K119-K20D3



All dimensions are in mm; tolerances according to ISO 2768 m-H

Interface

RPC-2.92 according to RPC-2.92 mechanically compatible with SMP according to

IEC 60169-35 RPC-3.50 and SMA MIL-STD 348A, Fig. 326

Documents

Application note

AN001 "Calibration Services"

Material and plating

Connector parts Center conductor Outer conductor RPC-2.92 Outer conductor SMP Dielectric

Material CuBe Stainless steel CuBe PS

Plating Gold, min. 1.27 µm, over nickel Passivated Gold, min. 1.27 µm, over nickel

35/09.14/6.2 Ш

Rosenberger Hochfrequenztechnik GmbH & Co. KG					
P.O.Box 1260	D-84526 Tittmoning	Germany			
www.rosenberger.de					

Tel. : +49 8684 18-0 Email : info@rosenberger.de

Technical Data Sheet

Rosenberger

Calibration Adaptor RPC-2.92 Jack – SMP Jack

02K119-K20D3

Electrical data

Frequency Return loss

DC to 40 GHz				
\geq 32 dB, DC to 12 GHz				
\geq 26 dB, 12 GHz to 26.5 GHz				
\geq 21 dB, 26.5 GHz to 40 GHz				

Mechanical data

Mechanical data		
		SMP
Mating cycles	≥ 500	\geq 100 if mating part is full detent
		\geq 500 if mating part is limited detent
		\geq 1000 if mating part is smooth bore
Maximum torque	1.70 Nm	
Recommended torque	0.90 Nm	
Engagement force		Full detent 68 N
		Limited detent 45 N
		Smooth bore 9 N
Disengagement force		Full detent 22 N
		Limited detent 9 N
		Smooth bore 2.2 N
Gauge	0.00 mm to 0.03 mm	0.00 mm to 0.05 mm

General standard definition

For proper operation the vector network analyzer (VNA) needs a model describing the electrical behaviour of this calibration standard. The different models, units, and terms used will depend on the VNA type and they will have to be entered into the VNA. All values are based on typical geometry and plating.

Offset Z_o / Impedance / Z_o Offset Delay Length (electrical) / Offset Length Offset Loss Loss 50 Ω 51.402 ps 15.41 mm 4.00 GΩ/s 0.0179 dB/ √GHz

Environmental data

Operating temperature range¹ Rated temperature range of use² Storage temperature range +20 °C to +26 °C 0 °C to +50 °C -40 °C to +85 °C

compliant

RoHS

¹ Temperature range over which these specifications are valid.

² This range is underneath and above the operating temperature range, within the open circuit is fully functional and could be used without damage.

Rosenberger Hochfrequenztechnik GmbH & Co. KG					
P.O.Box 1260	D-84526 Tittmoning	Germany			
www.rosenberger.de					

Tel. : +49 8684 18-0 Email : info@rosenberger.de

Technical Data Sheet

Rosenberger

Calibration Adaptor RPC-2.92 Jack – SMP Jack

02K119-K20D3

Declaration of calibration options

Factory Calibration

Standard delivery for this calibration standard includes a Factory Calibration. The Calibration Certificate issued reports individual calibration results, **traceable to Rosenberger standards**, national / international standards are not available. Model based standard definitions are reported in an Agilent/Keysight, Rohde & Schwarz and Anritsu compatible VNA format.

Accredited Calibration

Not available.

For further, more detailed information see application note AN001 on the Rosenberger homepage.

Calibration interval

Recommendation

12 months

Packing

Standard Weight 1 pce in box 3.3 g/pce

RF 35/09.14/6.2

, Ц

While the information has been carefully compiled to the best of our knowledge, nothing is intended as representation or warranty on our part and no statement herein shall be construed as recommendation to infringe existing patents. In the effort to improve our products, we reserve the right to make changes judged to be necessary.

Draft	Date	Approved	Date		Rev.	Engineering change number	Name	Date
Marcel Panicke	15/07/14	Herbert Babinger	08/12/15		b00	14-1492	M.Ruf	08/12/15
Rosenberger Hochfrequenztechnik GmbH & Co. KG				Tel.	Tel. : +49 8684 18-0		Page	
P.O.Box 1260 D-84526 Tittmoning Germany www.rosenberger.de			Ema	ail : info@rosenberger.de		3/3		