

Chipsmall Limited consists of a professional team with an average of over 10 year of expertise in the distribution of electronic components. Based in Hongkong, we have already established firm and mutual-benefit business relationships with customers from, Europe, America and south Asia, supplying obsolete and hard-to-find components to meet their specific needs.

With the principle of "Quality Parts, Customers Priority, Honest Operation, and Considerate Service", our business mainly focus on the distribution of electronic components. Line cards we deal with include Microchip, ALPS, ROHM, Xilinx, Pulse, ON, Everlight and Freescale. Main products comprise IC, Modules, Potentiometer, IC Socket, Relay, Connector. Our parts cover such applications as commercial, industrial, and automotives areas.

We are looking forward to setting up business relationship with you and hope to provide you with the best service and solution. Let us make a better world for our industry!



Contact us

Tel: +86-755-8981 8866 Fax: +86-755-8427 6832

Email & Skype: info@chipsmall.com Web: www.chipsmall.com

Address: A1208, Overseas Decoration Building, #122 Zhenhua RD., Futian, Shenzhen, China








Dieses Dokument ist urheberrechtlich geschützt • This document is protected by copyright • Rosenberger Hochfrequenztechnik GmbH & Co. KG

RF 35/09.14/6.2

Technical Data Sheet RPC-2.92 Calibration Adaptor Plug/Plug RPC-2.92 Calibration Adaptor Plug/Plug



All dimensions are in mm; tolerances according to ISO 2768 m-H

Interface

According to Mechanically compatible with

IEC 61169-35 RPC-3.50 and SMA

Documents

Application note

AN001 "Calibration Services"

Material and plating

Connector parts
Center conductor
Outer conductor
Coupling nut
Dielectric

Material

CuBe Stainless steel Stainless steel PS

Plating

Gold, min. 1.27 µm, over nickel Passivated Passivated

Rosenberger Hochfrequenztechnik GmbH & Co. KG P.O.Box 1260 D-84526 Tittmoning Germany www.rosenberger.de

Tel. : +49 8684 18-0 Email : info@rosenberger.de Page

1/3

Dieses Dokument ist urheberrechtlich geschützt • This document is protected by copyright • Rosenberger Hochfrequenztechnik GmbH & Co. KG

RF 35/09.14/6.2

Technical Data Sheet

Rosenberger

RPC-2.92

Calibration Adaptor Plug/Plug

02S121-S20S3

Electrical data

Frequency range Return loss DC to 40 GHz \geq 32 dB, DC to 4 GHz \geq 28 dB, 4 GHz to 18 GHz \geq 25 dB, 18 GHz to 40 GHz

Mechanical data

 $\begin{array}{ll} \text{Mating cycles} & \geq 500 \\ \text{Maximum torque} & 1.70 \text{ Nm} \\ \text{Recommended torque} & 0.90 \text{ Nm} \\ \end{array}$

Gauge 0.00 mm to 0.03 mm

General standard definitions

For proper operation the vector network analyzer (VNA) needs a model describing the electrical behaviour of this calibration standard. The different models, units, and terms used will depend on the VNA type and they will have to be entered into the VNA. All values are based on typical geometry and plating.

 $\begin{array}{lll} \text{Offset Z}_{\text{o}} \ / \ \text{Impedance} \ / \ Z_{\text{o}} & 50 \ \Omega \\ \text{Offset Delay} & 50.0346 \ \text{ps} \\ \text{Length (electrical)} \ / \ \text{Offset Length} & 15.00 \ \text{mm} \\ \text{Offset Loss} & 2.70 \ \text{G}\Omega/\text{s} \\ \text{Loss} & 0.0117 \ \text{dB}/\sqrt{\text{GHz}} \end{array}$

Environmental data

Operating temperature range¹ +20 °C to +26 °C Rated temperature range of use² 0 °C to +50 °C Storage temperature range -40 °C to +85 °C

RoHS compliant

Tel. : +49 8684 18-0

¹ Temperature range over which these specification are valid.

² This range is underneath and above the operating temperature range, within the calibration adaptor is fully functional and could be used without damage.

Technical Data Sheet RPC-2.92 Calibration Adaptor Plug/Plug RPC-2.92 Calibration Adaptor Plug/Plug

Declaration of calibration options

Factory Calibration

Standard delivery for this calibration standard includes a Factory Calibration. The Calibration Certificate issued reports individual calibration results, traceable to national / international standards. Model based standard definitions are reported in an Agilent/Keysight, Rohde & Schwarz and Anritsu compatible VNA format

Accredited Calibration

Optional this calibration standard can be delivered with an Accredited Calibration (DAkkS) having the highest confidence in the traceability. The DAkkS Calibration Certificate issued reports individual calibration results in a complex format, traceable to national / international standards. Model based standard definitions are reported in an Agilent/Keysight, Rohde & Schwarz and Anritsu compatible VNA format. The uncertainties are smaller than in a Factory Calibration.

For further, more detailed information see application note AN001 on the Rosenberger homepage.

Calibration interval				
Recommendation	12 months			
Packing				
Standard	1 pce in box			
Weight	6.1 g/pce			

While the information has been carefully compiled to the best of our knowledge, nothing is intended as representation or warranty on our part and no statement herein shall be construed as recommendation to infringe existing patents. In the effort to improve our products, we reserve the right to make changes judged to be necessary.

Draft	Date	Approved	Date		Date		Rev.	Engineering change number	Name		Date
Herbert Babinger	01.12.14	Markus Müller	03.11.16		f00	16-1390	Marion Striegler	(03.11.16		
		1 " 0 11100	140								

Rosenberger Hochfrequenztechnik GmbH & Co. KG P.O.Box 1260 D-84526 Tittmoning Germany www.rosenberger.de

Tel. : +49 8684 18-0 Email : info@rosenberger.de Page 3 / 3