



Chipsmall Limited consists of a professional team with an average of over 10 year of expertise in the distribution of electronic components. Based in Hongkong, we have already established firm and mutual-benefit business relationships with customers from,Europe,America and south Asia,supplying obsolete and hard-to-find components to meet their specific needs.

With the principle of “Quality Parts,Customers Priority,Honest Operation,and Considerate Service”,our business mainly focus on the distribution of electronic components. Line cards we deal with include Microchip,ALPS,ROHM,Xilinx,Pulse,ON,Everlight and Freescale. Main products comprise IC,Modules,Potentiometer,IC Socket,Relay,Connector.Our parts cover such applications as commercial,industrial, and automotives areas.

We are looking forward to setting up business relationship with you and hope to provide you with the best service and solution. Let us make a better world for our industry!



Contact us

Tel: +86-755-8981 8866 Fax: +86-755-8427 6832

Email & Skype: info@chipsmall.com Web: www.chipsmall.com

Address: A1208, Overseas Decoration Building, #122 Zhenhua RD., Futian, Shenzhen, China





Picture shows similar calibration kit

Contents

Device	Part number	QTY	Calibration Option ^a
20 dB Attenuator plug/jack	05AS122-K20S3	1	FC
40 dB Attenuator plug/jack	05AS122-K40S3	1	FC
Mismatch Airline plug/jack	05S102-K100	1	FC
Including 50 Ω Center Conductor	05S101-K100/20D	1	
Combi wrench	53W011-000	1	-

a. See "Declaration of calibration options" for explanation.

This Verification Kit is designed as an electrical verification standard. Reference measurement values (calibration results) for transmission and reflection are included. Connected to a calibrated VNA the actual measured transmission and reflection values can be compared to the reference measurement values and the quality of the VNA calibration can be evaluated.

Documentation

This kit is delivered with

- **USB-Stick**
With S2p data file of the reference measurement values. Calibration Certificate as PDF-file.
- **Kit Info Card**
Handling precautions.
- **Calibration Certificate**
Details see "Declaration of calibration options".
- **Operating Manual**

Electrical specifications

This specification covers electrical key values for the main items of the verification kit. Specific datasheets are available for each component among the part number.

Device	Frequency	Parameter	Specification
20 dB Attenuator	DC to ≤ 4 GHz > 4 GHz to ≤ 18 GHz DC to ≤ 4 GHz > 4 GHz to ≤ 18 GHz	Return Loss Return Loss Insertion Loss Insertion Loss	≥ 32 dB ≥ 26 dB 20 dB ± 0.30 dB 20 dB ± 0.50 dB
40 dB Attenuator	DC to ≤ 4 GHz > 4 GHz to ≤ 18 GHz DC to ≤ 4 GHz > 4 GHz to ≤ 18 GHz	Return Loss Return Loss Insertion Loss Insertion Loss	≥ 32 dB ≥ 23 dB 40 dB ± 0.50 dB 40 dB ± 1.00 dB
25 Ω Mismatch Airline		Outer conductor: Diameter Length Center Conductor: Diameter 50 Ω section Diameter 25 Ω section Length total Length 25 Ω section	7.000 mm nominal 100.00 mm nominal 3.040 mm nominal 4.615 mm nominal 100.00 mm nominal 75.00 mm nominal
50 Ω Airline Center Conductor		Diameter Length	3.040 mm nominal 100.00 mm nominal

Dieses Dokument ist urheberrechtlich geschützt • This document is protected by copyright • Rosenberger Hochfrequenztechnik GmbH & Co. KG

RF_35/05.10/6.1

Declaration of calibration options

Factory Calibration

Standard delivery for this verification standard includes a Factory Calibration. The Calibration Certificate issued reports individual calibration results, traceable to national / international standards. A S2p data file with the reference measurement values is included.

Accredited Calibration

Not available.

For further, more detailed information see application note AN001 on the Rosenberger homepage.

Calibration interval

Recommendation 12 months

Recommended accessories

- Rosenberger Test Port Adaptor
- Rosenberger Gauge Kit 05GK0KS-010
- Rosenberger VNA Test cable kit and Microwave Cable Assemblies

For further, more detailed information please visit our homepage www.rosenberger.com.

While the information has been carefully compiled to the best of our knowledge, nothing is intended as representation or warranty on our part and no statement herein shall be construed as recommendation to infringe existing patents. In the effort to improve our products, we reserve the right to make changes judged to be necessary.

Draft	Date	Approved	Date	Rev.	Engineering change number	Name	Date
Herbert Babinger	10.08.15	Martin Moder	12.08.15	a00	15-s228	A. Youmsi Mouafo	12.08.15

Rosenberger Hochfrequenztechnik GmbH & Co. KG P.O.Box 1260 D-84526 Tittmoning Germany www.rosenberger.de	Tel. : +49 8684 18-0 Fax : +49 8684 18-499 Email : info@rosenberger.de	Page 3 / 3
--	---	---------------