

Chipsmall Limited consists of a professional team with an average of over 10 year of expertise in the distribution of electronic components. Based in Hongkong, we have already established firm and mutual-benefit business relationships with customers from, Europe, America and south Asia, supplying obsolete and hard-to-find components to meet their specific needs.

With the principle of "Quality Parts, Customers Priority, Honest Operation, and Considerate Service", our business mainly focus on the distribution of electronic components. Line cards we deal with include Microchip, ALPS, ROHM, Xilinx, Pulse, ON, Everlight and Freescale. Main products comprise IC, Modules, Potentiometer, IC Socket, Relay, Connector. Our parts cover such applications as commercial, industrial, and automotives areas.

We are looking forward to setting up business relationship with you and hope to provide you with the best service and solution. Let us make a better world for our industry!



## Contact us

Tel: +86-755-8981 8866 Fax: +86-755-8427 6832

Email & Skype: info@chipsmall.com Web: www.chipsmall.com


Address: A1208, Overseas Decoration Building, #122 Zhenhua RD., Futian, Shenzhen, China







Technical Data Sheet		Rosenberger		
RPC-N 50 Ω	Calibration Load Jack	05K150-C10S3		



All dimensions are in mm; tolerances according to ISO 2768 m-H

Interface	
According to	IEC 61169-16

Documents	
Application note	AN001 "Calibration Services"

Material and plating		
Connector parts	Material	Plating
Center conductor	CuBe	Gold, min. 1.27 μm, over nickel
Outer conductor	Stainless steel	Passivated
Dielectric	PPE	
Substrate	$Al_2O_3$	

Tel. : +49 8684 18-0

Email: info@rosenberger.de

Rosenberger Hochfrequenztechnik GmbH & Co. KG P.O.Box 1260 D-84526 Tittmoning Germany www.rosenberger.de

Page

1/3

# Technical Data Sheet Rosenberger

RPC-N Calibration Load 50 Ω SK150-C10S3

#### Electrical data

Frequency range DC to 18 GHz

Return loss  $\geq$  45 dB, DC to 4 GHz

≥ 32 dB, 4 GHz to 8 GHz ≥ 30 dB, 8 GHz to 18 GHz

DC Resistance 50  $\Omega \pm 0.25 \Omega$ 

Power handling ≤ 1 W

#### Mechanical data

 $\begin{array}{ll} \text{Mating cycles} & \geq 500 \\ \text{Maximum torque} & 1.70 \text{ Nm} \\ \text{Recommended torque} & 1.10 \text{ Nm} \\ \end{array}$ 

Gauge 5.22 mm to 5.26 mm

#### **General standard definitions**

For proper operation the vector network analyzer (VNA) needs a model describing the electrical behaviour of this calibration standard. The different models, units, and terms used will depend on the VNA type and they will have to be entered into the VNA. All values are based on typical geometry and plating.

 $\begin{array}{ll} \text{Offset $Z_{\circ}$ / Impedance / $Z_{\circ}$} & 50 \ \Omega \\ \text{Offset Delay} & 0.0000 \ \text{ps} \\ \text{Length (electrical) / Offset Length} & 0.00 \ \text{mm} \\ \text{Offset Loss} & 0.00 \ \text{G}\Omega/\text{s} \\ \end{array}$ 

Loss  $0.0000 \, dB/\sqrt{GHz}$ 

#### **Environmental data**

Operating temperature range<sup>1</sup> +20 °C to +26 °C
Rated temperature range of use<sup>2</sup> 0 °C to +50 °C
Storage temperature range -40 °C to +85 °C

RoHS compliant

Rosenberger Hochfrequenztechnik GmbH & Co. KG P.O.Box 1260 D-84526 Tittmoning Germany www.rosenberger.de

Tel. : +49 8684 18-0 Email : info@rosenberger.de Page

2/3

<sup>&</sup>lt;sup>1</sup> Temperature range over which these specification are valid.

<sup>&</sup>lt;sup>2</sup> This range is underneath and above the operating temperature range, within the calibration load is fully functional and could be used without damage.

Technical Data Sheet		Rosenberger		
RPC-N 50 Ω	Calibration Load Jack	05K150-C10S3		

#### **Declaration of calibration options**

### **Factory Calibration**

Standard delivery for this calibration standard includes a Factory Calibration. The Calibration Certificate issued reports individual calibration results, traceable to national / international standards. Model based standard definitions are reported in an Agilent/Keysight, Rohde & Schwarz and Anritsu compatible VNA format.

#### **Accredited Calibration**

Optional this calibration standard can be delivered with an Accredited Calibration (DAkkS) having the highest confidence in the traceability. The DAkkS Calibration Certificate issued reports individual calibration results in a complex format, traceable to national / international standards. Model based standard definitions are reported in an Agilent/Keysight, Rohde & Schwarz and Anritsu compatible VNA format as well as in a dense data set needed for data based standard definitions. The uncertainties are smaller than in a Factory Calibration.

For further, more detailed information see application note AN001 on the Rosenberger homepage.

### **Calibration** interval

Recommendation

12 months

#### **Packing**

Standard Weight 1 pce in box 45.0 g/pce

While the information has been carefully compiled to the best of our knowledge, nothing is intended as representation or warranty on our part and no statement herein shall be construed as recommendation to infringe existing patents. In the effort to improve our products, we reserve the right to make changes judged to be necessary.

Draft	Date	Approved	Date	Rev.	Engineering change number	Name	Date
Herbert Babinger	14.02.05	Lars Ramtke	19.03.15	e00	15-0329	Marion Striegler	19.03.15

Rosenberger Hochfrequenztechnik GmbH & Co. KG P.O.Box 1260 D-84526 Tittmoning Germany www.rosenberger.de

Tel. : +49 8684 18-0 Email : info@rosenberger.de Page

3/3