imall

Chipsmall Limited consists of a professional team with an average of over 10 year of expertise in the distribution of electronic components. Based in Hongkong, we have already established firm and mutual-benefit business relationships with customers from, Europe, America and south Asia, supplying obsolete and hard-to-find components to meet their specific needs.

With the principle of "Quality Parts, Customers Priority, Honest Operation, and Considerate Service", our business mainly focus on the distribution of electronic components. Line cards we deal with include Microchip, ALPS, ROHM, Xilinx, Pulse, ON, Everlight and Freescale. Main products comprise IC, Modules, Potentiometer, IC Socket, Relay, Connector. Our parts cover such applications as commercial, industrial, and automotives areas.

We are looking forward to setting up business relationship with you and hope to provide you with the best service and solution. Let us make a better world for our industry!



Contact us

Tel: +86-755-8981 8866 Fax: +86-755-8427 6832 Email & Skype: info@chipsmall.com Web: www.chipsmall.com Address: A1208, Overseas Decoration Building, #122 Zhenhua RD., Futian, Shenzhen, China



Technic	al Data Sheet		Rosenberg				
RPC-1.85	C-1.85 Calibration Adaptor 08K121-K20S3						
serial ident. no.							
Interface According to Mechanicall		IEC 61169-32 RPC-2.40					
Document Application		AN001 "Calibrat	ion Services"				
Material an Connector Center cond Outer cond Dielectric	parts ductor uctor	Material Beryllium coppe Stainless steel PS	Plating r Gold, min. 1.27 μm, over nickel Passivated				
Decemberger II	ochfrequenztechnik GmbH & Co. KG	Tel.	: +49 8684 18-0	Page			

osenberger Hochfrequenztechnik GmbH & Co. KG	Tel. : +49 8684 18-0	Page
.O.Box 1260 D-84526 Tittmoning Germany ww.rosenberger.de	Fax : +49 8684 18-499 Email : <u>info@rosenberger.de</u>	1/3

Rosenberger

RPC-1.85

Calibration Adaptor Jack/Jack

08K121-K20S3

Electrical data Frequency range

Return loss

DC to 70 GHz \geq 28 dB, DC to 4 GHz \geq 20 dB, 4 GHz to 40 GHz \geq 17 dB, 40 GHz to 70 GHz

Mechanical data

Mating cycles Maximum torque Recommended torque Gauge ≥ 500 1.65 Nm 0.90 Nm 0.00 mm to 0.03 mm

General standard definitions

For proper operation the vector network analyzer (VNA) needs a model describing the electrical behaviour of this calibration standard. The different models, units, and terms used will depend on the VNA type and they will have to be entered into the VNA. All values are based on typical geometry and plating.

Offset Z_o / Impedance / Z_o Offset Delay Length (electrical) / Offset Length Offset Loss Loss 50 Ω 61.042 ps 18.3 mm 4.00 GΩ/s 0.0212 dB/ √GHz

Environmental data

Operating temperature range¹ Rated temperature range of use² Storage temperature range +20 °C to +26 °C 0 °C to +50 °C -40 °C to +85 °C

RoHS

compliant

¹ Temperature range over which these specifications are valid.

² This range is underneath and above the operating temperature range, within the calibration adaptor is fully functional and could be used without damage.

Rosenberger Hochfrequenztechnik GmbH & Co. KG	Tel. : +49 8684 18-0	Page
P.O.Box 1260 D-84526 Tittmoning Germany	Fax : +49 8684 18-499	_
www.rosenberger.de	Email : info@rosenberger.de	2/3

Technica	al Data Sheet	Rosenberger
RPC-1.85	Calibration Adaptor Jack/Jack	08K121-K20S3
Factory Ca Standard d issued repo standard d format. Accredited Not availab	elivery for this calibration standorts individual calibration result efinitions are reported in an Ag d Calibration ble.	dard includes a Factory Calibration. The Calibration Certificate ts, traceable to national / international standards. Model based gilent/Keysight, Rohde & Schwarz and Anritsu compatible VNA
Calibration Recommen		12 months
Packing Standard Weight		1 pce in box 6.4 g/pce

While the information has been carefully compiled to the best of our knowledge, nothing is intended as representation or warranty on our part and no statement herein shall be construed as recommendation to infringe existing patents. In the effort to improve our products, we reserve the right to make changes judged to be necessary.

					-		1			
Draft	Date	Approved	Date		Rev.	Engineering change number	Name		Date	
Herbert Babinger	25/08/04	Martin Moder	24.03.15	e00		14-1492	Herbert Babinger		24.03.15	
Rosenberger Hochfrequenztechnik GmbH & Co. KG P.O.Box 1260 D-84526 Tittmoning Germany				Tel. Fax	: +49 8684 18-0 : +49 8684 18-499			Page		
www.rosenberger.de					ail : <u>info@rosenberger.de</u>			3/3		