

Chipsmall Limited consists of a professional team with an average of over 10 year of expertise in the distribution of electronic components. Based in Hongkong, we have already established firm and mutual-benefit business relationships with customers from, Europe, America and south Asia, supplying obsolete and hard-to-find components to meet their specific needs.

With the principle of "Quality Parts, Customers Priority, Honest Operation, and Considerate Service", our business mainly focus on the distribution of electronic components. Line cards we deal with include Microchip, ALPS, ROHM, Xilinx, Pulse, ON, Everlight and Freescale. Main products comprise IC, Modules, Potentiometer, IC Socket, Relay, Connector. Our parts cover such applications as commercial, industrial, and automotives areas.

We are looking forward to setting up business relationship with you and hope to provide you with the best service and solution. Let us make a better world for our industry!



Contact us

Tel: +86-755-8981 8866 Fax: +86-755-8427 6832

Email & Skype: info@chipsmall.com Web: www.chipsmall.com


Address: A1208, Overseas Decoration Building, #122 Zhenhua RD., Futian, Shenzhen, China







Technica	al Data Sheet	Rosenberger			
RPC-1.85	Airline Plug / Jack	08S101-K013			



All dimensions are in mm; tolerances according to ISO 2768 m-H

Interface

According to Mechanically compatible with IEC 61169-32 RPC-2.40

Documents

Application note

AN001 "Calibration Services"

Material and plating

Connector parts Center conductor Outer conductor Coupling nut

Material

CuBe Gold, min. 1.27 µm, over chemical nickel Gold, min. 1.27 µm, over chemical nickel **Brass**

Passivated Stainless steel

Rosenberger Hochfrequenztechnik GmbH & Co. KG P.O.Box 1260 D-84526 Tittmoning Germany www.rosenberger.de

Page

Tel. : +49 8684 18-0 Email: info@rosenberger.de

1/3

Dieses Dokument ist urheberrechtlich geschützt • This document is protected by copyright • Rosenberger Hochfrequenztechnik GmbH & Co.

RF 35/09.14/6.2

Technical Data Sheet Rosenberger

RPC-1.85

Airline Plug / Jack

08S101-K013

Electrical data

Insertion loss

≤ 0.18 dB at 70 GHz

Mechanical data

Mating cycles≥ 500Maximum torque1.65 NmRecommended torque0.90 Nm

Airline dimensions at 23 °C:

- Diameter outer conductor

- Diameter inner conductor

- Length outer conductor

- Length inner conductor

- Length difference ≤ 0.04 mm

(outer conductor – inner conductor)

Calculated data (non warranted)

Lossless characteristic impedance Return loss² 50 Ω ± 1.00 Ω ≥ 34 dB, 0.5 GHz to 4 GHz ≥ 26 dB, 4 GHz to 40 GHz ≥ 24 dB, 40 GHz to 70 GHz

- 1. The lossless characteristic impedance is calculated from the specified diameters of the inner and outer conductor.
- 2. The return loss is calculated from the characteristic impedance, the skin depth and the connector interface.

General standard definitions

For proper work the vector network analyzer (VNA) used needs a model describing the electrical behaviour of this calibration standard. Depending on the VNA type different models, units and terms are used and have to be entered into the VNA. All values are based on typical geometry and plating.

 $\begin{array}{lll} \text{- Offset Z_{\circ} / Impedance / Z_{\circ}} & 50 \ \Omega \\ \text{- Offset Delay} & 43.4108 \ ps \\ \text{- Length (electrical) / Offset Length} & 13.014 \ mm \\ \text{- Offset Loss} & 4.50 \ G\Omega/s \end{array}$

- Loss $0.0170 \text{ dB}/\sqrt{\text{GHz}}$

Environmental data

Operating temperature range Storage temperature range RoHS

+20 °C to +26 °C 0 °C to +50 °C compliant

This range is a recommendation. However, the airline can be used in a wider range.
 Any temperature change from 23 °C results in dimensional changes.

Rosenberger Hochfrequenztechnik GmbH & Co. KG P.O.Box 1260 D-84526 Tittmoning Germany www.rosenberger.de

Page

Tel. : +49 8684 18-0 Email : info@rosenberger.de

Technical Data Sheet Rosenberger

RPC-1.85 | Airline Plug / Jack

08S101-K013

Declaration of calibration options

Factory Calibration

Standard delivery for this calibration standard includes a Factory Calibration. The Calibration Certificate issued reports individual mechanical calibration results, traceable to national / international standards. Model based standard definitions are reported in an Agilent/Keysight, Rohde & Schwarz and Anritsu compatible VNA format.

Accredited Calibration

Not available.

For further, more detailed information see application note AN001 on the Rosenberger homepage.

Calibration interval

Recommendation

12 months

Packing

Standard Weight

Center conductor

1 pce in box 4.5 g/pce

loose in an acrylic glass tube

While the information has been carefully compiled to the best of our knowledge, nothing is intended as representation or warranty on our part and no statement herein shall be construed as recommendation to infringe existing patents. In the effort to improve our products, we reserve the right to make changes judged to be necessary.

Herbert Babinger 26.08.04 Martin Moder 24.03.15 f00 14-1492 Herbert Babinger 24.03.15	Draft	Date	Approved	Date	Rev.	Engineering change number	Name	Date
	Herbert Babinger	26.08.04	Martin Moder	24.03.15	f00	14-1492	Herbert Babinger	24.03.15

Rosenberger Hochfrequenztechnik GmbH & Co. KG P.O.Box 1260 D-84526 Tittmoning Germany www.rosenberger.de

Tel. : +49 8684 18-0 Email : info@rosenberger.de Page 3/3