mail

Chipsmall Limited consists of a professional team with an average of over 10 year of expertise in the distribution of electronic components. Based in Hongkong, we have already established firm and mutual-benefit business relationships with customers from, Europe, America and south Asia, supplying obsolete and hard-to-find components to meet their specific needs.

With the principle of "Quality Parts, Customers Priority, Honest Operation, and Considerate Service", our business mainly focus on the distribution of electronic components. Line cards we deal with include Microchip, ALPS, ROHM, Xilinx, Pulse, ON, Everlight and Freescale. Main products comprise IC, Modules, Potentiometer, IC Socket, Relay, Connector. Our parts cover such applications as commercial, industrial, and automotives areas.

We are looking forward to setting up business relationship with you and hope to provide you with the best service and solution. Let us make a better world for our industry!



Contact us

Tel: +86-755-8981 8866 Fax: +86-755-8427 6832 Email & Skype: info@chipsmall.com Web: www.chipsmall.com Address: A1208, Overseas Decoration Building, #122 Zhenhua RD., Futian, Shenzhen, China




THESE DRAWINGS ARE THE PROPERTY OF QUALTEK ELECTRONICS AND NOT TO BE REPRODUCED FOR OR COMMUNICATED TO A THIRD PARTY WITHOUT THE EXPRESS WRITTEN PERMISSION OF AN OFFICER OF QUALTEK ELECTRONICS.

0

• 🔍

(2)

	REVISION		
REV.	DESCRIPTION	DATE	APPROVED
В	RoHS Compliancy	06-01-06	DP



CONTROLLED

NOTES:

1.	CORDAGE: SJT, 18AWG / 3C , UNSHIELDED,
	NA COLOR CODE, TEMP. RATING 60°C,
	RATING: 125V, 10A, JACKET COLOR: GRAY
2.	PLUG: NEMA 5-15P HOSPITAL GRADE (GRAY)
3.	CONNECTOR: IEC 60320 C-13 RIGHT ANGLED (GRAY)
4.	APPROVALS: UL, CSA
-	

- 5. ROHS COMPLIANT
- 6. CAUTION LABEL

	Dim range	Tolerance +/- (mm)	Qualtek Electronics Corp. PPC DIVISION			
	(mm)		Part Number:	240008-06	RoHS Compliant	
	0-1	0.15	Description: POWER CORD HOSPITAL GRADE			
	1-6	0.20	POWER CORD HOSPITAL GRADE			
	6-18	0.30				
ľ	18-50	0.50		UNIT: in	REV. B	
	50-120	0.70	Drawn / Date	Checked / Date	Approved / Date	
	120-250	0.90	B.Zhang	02-20-02	JJH / 02-20-02	
	250-500	1.20	02-20-02		JJ17/ 02-20-02	

96″±

CAUTION

GROUNDING RELIABILITY CAN ONLY BE ACHIEVED WHEN THE EQUIPMENT IS CONNECTED TO AN EQUIVALENT RECEPTACLE MARKED HOSPITAL ONLY OR HOSPITAL GRADE