Chipsmall Limited consists of a professional team with an average of over 10 year of expertise in the distribution of electronic components. Based in Hongkong, we have already established firm and mutual-benefit business relationships with customers from,Europe,America and south Asia,supplying obsolete and hard-to-find components to meet their specific needs.

With the principle of "Quality Parts,Customers Priority,Honest Operation,and Considerate Service",our business mainly focus on the distribution of electronic components. Line cards we deal with include Microchip,ALPS,ROHM,Xilinx,Pulse,ON,Everlight and Freescale. Main products comprise IC,Modules,Potentiometer,IC Socket,Relay,Connector.Our parts cover such applications as commercial,industrial, and automotives areas.

We are looking forward to setting up business relationship with you and hope to provide you with the best service and solution. Let us make a better world for our industry!

Contact us

Tel: +86-755-8981 8866 Fax: +86-755-8427 6832
Email & Skype: info@chipsmall.com Web: www.chipsmall.com
Address: A1208, Overseas Decoration Building, #122 Zhenhua RD., Futian, Shenzhen, China

Digital Sound Level Meter

Analog bargraph with 50dB range
Updates every 40ms for viewing trends

Features:

- ±2dB accuracy with 0.1dB resolution
- A&C weighting
- AC analog output
- Record Max/Min values over time
- Auto power off and Max Hold functions
- Utilizes 0.5"(12.7mm) condenser microphone
- Tripod mountable
- 40 to 130dB measuring range
- Fast/slow response time
- Tripod mountable
- Complete with 4 AAA batteries and microphone windscreen

Applications:

- Monitoring workplace machinery
- Installing audio and alarm systems
- Product noise certification and reduction

Specifications	
Display Counts	2000 count LCD
Range	40 to 130dB
Basic Accuracy	±2dB
Condenser Microphone	0.5" (12.7mm)
Analog Output	AC
Dimensions	9 x 2.2 x 1.7"(230 x 57 x 44mm)
Weight	5.6 oz (160g)

Ordering Information:

407730....................Digital Sound Level Meter
407730-NIST ....407730 with NIST Certificate

CE

9/27/13 - R1