

Chipsmall Limited consists of a professional team with an average of over 10 year of expertise in the distribution of electronic components. Based in Hongkong, we have already established firm and mutual-benefit business relationships with customers from, Europe, America and south Asia, supplying obsolete and hard-to-find components to meet their specific needs.

With the principle of "Quality Parts, Customers Priority, Honest Operation, and Considerate Service", our business mainly focus on the distribution of electronic components. Line cards we deal with include Microchip, ALPS, ROHM, Xilinx, Pulse, ON, Everlight and Freescale. Main products comprise IC, Modules, Potentiometer, IC Socket, Relay, Connector. Our parts cover such applications as commercial, industrial, and automotives areas.

We are looking forward to setting up business relationship with you and hope to provide you with the best service and solution. Let us make a better world for our industry!



Contact us

Tel: +86-755-8981 8866 Fax: +86-755-8427 6832

Email & Skype: info@chipsmall.com Web: www.chipsmall.com


Address: A1208, Overseas Decoration Building, #122 Zhenhua RD., Futian, Shenzhen, China







Technic	al Data Sheet	Rosenberger			
BNC 50 Ω	Open Circuit Plug	51S12L-000S3			



All dimensions are in mm; tolerances according to ISO 2768 m-H

Interface

According to

IEC 60169-8, MIL-PRF-39012, CECC 22120

Documents

Application note

AN001 "Calibration Services"

Material and plating

Connector parts

Center conductor Outer conductor Dielectric

Material

CuBe Stainless steel PTFE, PS

Plating

Gold, min. 1.27 μm , over nickel Passivated

Rosenberger Hochfrequenztechnik GmbH & Co. KG P.O.Box 1260 D-84526 Tittmoning Germany www.rosenberger.de Tel. : +49 8684 18-0 Fax : +49 8684 18-1499 Email : info@rosenberger.de Page

1/3

Ā Dieses Dokument ist urheberrechtlich geschützt ● This document is protected by copyright ● Rosenberger Hochfrequenztechnik GmbH & Co.

F 35/05.10/6.0

Technical Data Sheet Rosenberger

BNC Open Circuit O Plug O 51S12L-000S3

Electrical data

Frequency range DC to 4 GHz

Return loss \leq 0.20 dB, DC to 2 GHz

≤ 0.50 dB, 2 GHz to 4 GHz

Error from nominal phase¹ $\leq 3.0^{\circ}$, DC to 2 GHz

 \leq 5.0°, 2 GHz to 4 GHz

Mechanical data

Mating cycles ≥ 50

Gauge 5.31 mm to 5.38 mm

General standard definitions

For proper operation the vector network analyzer (VNA) needs a model describing the electrical behaviour of this calibration standard. The different models, units, and terms used will depend on the VNA type and they will have to be entered into the VNA. All values are based on typical geometry and plating.

 $\begin{array}{ll} \text{Offset Z}_{\text{o}} \ / \ \text{Impedance} \ / \ Z_{\text{o}} & 50 \ \Omega \\ \text{Offset Delay} & 70.448 \ \text{ps} \\ \text{Length (electrical)} \ / \ \text{Offset Length} & 21.12 \ \text{mm} \\ \text{Offset Loss} & 1.2 \ \text{G}\Omega / \text{s} \end{array}$

Loss $0.0147 \, dB / \sqrt{GHz}$

Fringing Capacitances²

Environmental data

Operating temperature range³ +20 °C to +26 °C
Rated temperature range of use⁴ 0 °C to +50 °C
Storage temperature range -40 °C to +85 °C

RoHS compliant

¹ The nominal phase is defined by the Offset Delay, the Offset Loss and the Fringing Capacitances

² Fringing Capacitances are determined individually for each Open Circuit and are documented in a Calibration Certificate.

³ Temperature range over which these specifications are valid.

⁴ This range is underneath and above the operating temperature range, within the open circuit is fully functional and could be used without damage.

www.rosenberger.de

Declaration of calibration options

Factory Calibration

Standard delivery for this calibration standard includes a Factory Calibration. The Calibration Certificate issued reports individual calibration results, **traceable to Rosenberger standards**, national / international standards are not available. Model based standard definitions are individually optimized and reported in an Agilent/Keysight, Rohde & Schwarz and Anritsu compatible VNA format.

Accredited Calibration

Not available.

For further, more detailed information see application note AN001 on the Rosenberger homepage.

Calibration interval

Recommendation 12 months

Packing

Standard Weight

1 pce in box 47 g/pce

While the information has been carefully compiled to the best of our knowledge, nothing is intended as representation or warranty on our part and no statement herein shall be construed as recommendation to infringe existing patents. In the effort to improve our products, we reserve the right to make changes judged to be necessary.

L	Draft	Date	Approved	Date	Date R		Engineering change number	Name		Date
I	Herbert Babinger	28.01.05	Martin Moder	09.07.15		e00	14-1492	Manfred Ruf		09.07.15
Rosenberger Hochfrequenztechnik GmbH & Co. KG Tel. : +49 8684 18-0										Page
P.O.Box 1260 D-84526 Tittmoning Germany							: +49 8684 18-1499			

Email: info@rosenberger.de

3/3