



Chipsmall Limited consists of a professional team with an average of over 10 year of expertise in the distribution of electronic components. Based in Hongkong, we have already established firm and mutual-benefit business relationships with customers from,Europe,America and south Asia,supplying obsolete and hard-to-find components to meet their specific needs.

With the principle of “Quality Parts,Customers Priority,Honest Operation,and Considerate Service”,our business mainly focus on the distribution of electronic components. Line cards we deal with include Microchip,ALPS,ROHM,Xilinx,Pulse,ON,Everlight and Freescale. Main products comprise IC,Modules,Potentiometer,IC Socket,Relay,Connector.Our parts cover such applications as commercial,industrial, and automotives areas.

We are looking forward to setting up business relationship with you and hope to provide you with the best service and solution. Let us make a better world for our industry!



Contact us

Tel: +86-755-8981 8866 Fax: +86-755-8427 6832

Email & Skype: info@chipsmall.com Web: www.chipsmall.com

Address: A1208, Overseas Decoration Building, #122 Zhenhua RD., Futian, Shenzhen, China





Picture shows similar calibration kit

Usage

Please note: This Calibration Kit includes a set of adaptors only. The full functionality is only given with an additional RPC-N calibration kit, details see necessary accessory.

Please see application note AN002 "S-Parameter Measurements with Non-Metrology Connector Interfaces" on the Rosenberger homepage for further, more detailed information.

Contents

| Device | Part number | Quantity | Calibration Option ^a |
|--------------------------------------|--------------|----------|---------------------------------|
| Adaptor RPC-N plug to 4.3-10 plug | 05S164-S20S3 | 2 | FC |
| Adaptor RPC-N jack to 4.3-10 jack | 05K164-K20S3 | 2 | FC |

a. See "Declaration of calibration options" for explanation.

Necessary accessory

- Rosenberger Calibration Kit RPC-N Full Version 05CK001-150
or
- Rosenberger Calibration Kit RPC-N Industrial Version 05CK010-150
or
- Equivalent OSLT Type N 50 Ohm Calibration Kit

Documentation

This kit is delivered with

- **USB-Stick**
Standard Definitions as data files for Vector Network Analyzer Family PNA (Keysight/Agilent).
Calibration Certificate as PDF-file.
- **Standard Definitions Cards**
Printed Standard Definitions that can be used on nearly all Vector Network Analyzers.
- **Kit Info Card**
Handling precautions and information for installing Standard Definitions on a Vector Network Analyzer.
- **Calibration Certificate**
Details see "Declaration of calibration options"
- **Operating Manual**

Electrical specifications

This specification covers electrical key values for the main calibration standards of the calibration kit. Specific datasheets are available for each component among the part number.

| Calibration standard | Frequency | Parameter | Specification |
|----------------------|---------------------|-------------|---------------|
| Calibration adaptors | DC to ≤ 4 GHz | Return Loss | ≥ 34 dB |
| | > 4 GHz to ≤ 6 GHz | | ≥ 30 dB |
| | > 6 GHz to ≤ 12 GHz | | ≥ 25 dB |

Declaration of calibration options

Factory Calibration

Standard delivery for this kit includes a Factory Calibration. The Calibration Certificate issued reports individual calibration results, **traceable to Rosenberger standards**, national / international standards are not available. Model based standard definitions of the calibration standards are reported in an Agilent/Keysight, Rohde & Schwarz and Anritsu compatible VNA format.

Accredited Calibration

Not available.

For further, more detailed information see application note AN001 on the Rosenberger homepage.

Calibration interval

Recommendation 12 months

Recommended accessories

- Rosenberger Test Port Adaptor
- Rosenberger VNA Test cable kit and Microwave Cable Assemblies

For further, more detailed information please visit our homepage www.rosenberger.com.

While the information has been carefully compiled to the best of our knowledge, nothing is intended as representation or warranty on our part and no statement herein shall be construed as recommendation to infringe existing patents. In the effort to improve our products, we reserve the right to make changes judged to be necessary.

| | | | | | | | |
|-----------|----------|------------|----------|------|---------------------------|-------|----------|
| Draft | Date | Approved | Date | Rev. | Engineering change number | Name | Date |
| F. Reiner | 12.05.16 | H.Babinger | 12.01.18 | d00 | 17-2178 | M.Ruf | 12.01.18 |