

Chipsmall Limited consists of a professional team with an average of over 10 year of expertise in the distribution of electronic components. Based in Hongkong, we have already established firm and mutual-benefit business relationships with customers from, Europe, America and south Asia, supplying obsolete and hard-to-find components to meet their specific needs.

With the principle of "Quality Parts, Customers Priority, Honest Operation, and Considerate Service", our business mainly focus on the distribution of electronic components. Line cards we deal with include Microchip, ALPS, ROHM, Xilinx, Pulse, ON, Everlight and Freescale. Main products comprise IC, Modules, Potentiometer, IC Socket, Relay, Connector. Our parts cover such applications as commercial, industrial, and automotives areas.

We are looking forward to setting up business relationship with you and hope to provide you with the best service and solution. Let us make a better world for our industry!



# Contact us

Tel: +86-755-8981 8866 Fax: +86-755-8427 6832

Email & Skype: info@chipsmall.com Web: www.chipsmall.com

Address: A1208, Overseas Decoration Building, #122 Zhenhua RD., Futian, Shenzhen, China







# Dieses Dokument ist urheberrechtlich geschützt • This document is protected by copyright • Rosenberger Hochfrequenztechnik GmbH & Co. KG


# RF 35/09.14/6.2

# **Technical Data Sheet**

# Rosenberger

Calibration Adaptor RPC-N 75  $\Omega$  Plug – BNC 75  $\Omega$  Jack

P5S171-K2AS3



All dimensions are in mm; tolerances according to ISO 2768 m-H

# Interface

RPC-N 75  $\Omega$  according to BNC 75  $\Omega$  according to

IEC 61169-16

IEC 61169-8, MIL-PRF-39012, CECC 22120

# **Documents**

Application note

AN001 "Calibration Services"

# Material and plating

# Connector parts

Center conductor Outer conductor Coupling nut Dielectric

# Material Plating

CuBe Gold, min. 1.27 µm, over nickel Stainless steel Passivated Passivated PS

Rosenberger Hochfrequenztechnik GmbH & Co. KG P.O.Box 1260 D-84526 Tittmoning Germany www.rosenberger.de

Tel.: +49 8684 18-0 Email: info@rosenberger.de Page

1/3

# ΑĞ Dieses Dokument ist urheberrechtlich geschützt • This document is protected by copyright • Rosenberger Hochfrequenztechnik GmbH & Co.

# 35/09.14/6.2 胎

# **Technical Data Sheet**

# Rosenberger

Calibration Adaptor RPC-N 75 Ω Plug – BNC 75 Ω Jack

P5S171-K2AS3

# Electrical data

Frequency Return loss

DC to 12 GHz ≥ 30 dB, DC to 4 GHz ≥ 25 dB, 4 GHz to 8 GHz ≥ 15 dB, 8 GHz to 12 GHz

### Mechanical data

Mating cycles

≥ 500

RPC-N 75 Ω BNC 75 Ω

Maximum torque 1.70 Nm Recommended torque 1.10 Nm

Gauge 5.28 mm to 5.36 mm 5.21 mm to 5.28 mm

# General standard definition

For proper operation the vector network analyzer (VNA) needs a model describing the electrical behaviour of this calibration standard. The different models, units, and terms used will depend on the VNA type and they will have to be entered into the VNA. All values are based on typical geometry and plating.

Offset Z<sub>o</sub> / Impedance / Z<sub>o</sub> Offset Delay 135.761 ps Length (electrical) / Offset Length 40.70 mm Offset Loss  $1.2 \,\mathrm{G}\Omega/\mathrm{s}$ 

0.0094 dB/√GHz Loss

# Environmental data

Operating temperature range +20 °C to +26 °C Rated temperature range of use<sup>2</sup> 0 °C to +50 °C Storage temperature range -40 °C to +85 °C

RoHS compliant

Rosenberger Hochfrequenztechnik GmbH & Co. KG P.O.Box 1260 D-84526 Tittmoning Germany www.rosenberger.de

Tel. : +49 8684 18-0

Email: info@rosenberger.de

Page

2/3

<sup>&</sup>lt;sup>1</sup> Temperature range over which these specifications are valid.

<sup>&</sup>lt;sup>2</sup> This range is underneath and above the operating temperature range, within the open circuit is fully functional and could be used without damage.

# Technical Data Sheet Rosenberger

Calibration Adaptor RPC-N 75  $\Omega$  Plug – BNC 75  $\Omega$  Jack

P5S171-K2AS3

# Declaration of calibration options

# **Factory Calibration**

Standard delivery for this calibration standard includes a Factory Calibration. The Calibration Certificate issued reports individual calibration results, **traceable to Rosenberger standards**, national / international standards are not available. Model based standard definitions are reported in an Agilent/Keysight, Rohde & Schwarz and Anritsu compatible VNA format.

### **Accredited Calibration**

Not available.

For further, more detailed information see application note AN001 on the Rosenberger homepage.

### Calibration interval

Recommendation

12 months

# **Packing**

Standard Weight 1 pce in box 40 g/pce

While the information has been carefully compiled to the best of our knowledge, nothing is intended as representation or warranty on our part and no statement herein shall be construed as recommendation to infringe existing patents. In the effort to improve our products, we reserve the right to make changes judged to be necessary.

Marcel Panicke 17.04.14	Martin Moder	24.02.15	d00	14-1492	Herbert Babinger	24.02.15

Rosenberger Hochfrequenztechnik GmbH & Co. KG P.O.Box 1260 D-84526 Tittmoning Germany www.rosenberger.de

Tel. : +49 8684 18-0 Email : info@rosenberger.de Page

3/3